



CÓDIGO:
ZP-0155

FICHA TÉCNICA

(Paquete de 50 piezas) Tapaboca reutilizable KN95. Mascarilla para mas de un uso con filtro triple. Nivel de proteccion KN95 que filtra el 95% de los virus y bacterias mismo nivel de proteccion que NIOSH N95

Based on this comparison, it is reasonable to consider China KN95, AS/NZ P2, Korea 1st Class, and Japan DS FF R3 as "equivalent" to US NIOSH N95 and European FFP2 respirators for filtering non-oil-based particles such as those resulting from wildfires, PM 2.5 air pollution, volcanic eruptions, or bioaerosols (e.g. viruses). However, prior to selecting a respirator, users should consult their local respiratory protection regulations and requirements or check with their local public health authorities for selection guidance.

Certification/Class (Standard)	N95 (NIOSH 42CFR 8430 FFRs)	FFP2 (EN 149:2001)	KN95 (GB2626-2006)	P2 (AS/NZ 3760:2012)	Korea 1st Class (KMOEL 2017-64)	DS (Japan JMAHLW-Notification 214, 2018)
Filter performance - (based on a 0.3 µm efficient)	≥ 95%	≥ 94%	≥ 95%	≥ 94%	≥ 94%	≥ 95%
Treatment	NaCl	NaCl and paraformal	NaCl	NaCl	NaCl and paraformal	NaCl
Flow rate	85 L/min	85 L/min	85 L/min	85 L/min	85 L/min	85 L/min
Total inward leakage (TIL) - based on human subjects wearing breathing apparatus	N/A	≤ 5% leakage (arithmetic mean)	≤ 5% leakage (arithmetic mean)	≤ 5% leakage (individual and arithmetic mean)	≤ 5% leakage (arithmetic mean)	Inward leakage measured and included in use instructions
Inhalation resistance - max pressure drop	≤ 243 Pa	≤ 70 Pa (at 30 L/min) + 243 Pa (at 95 L/min) + 250 Pa (snapfit)	≤ 250 Pa	≤ 70 Pa (at 30 L/min) + 240 Pa (at 95 L/min)	≤ 70 Pa (at 30 L/min) + 240 Pa (at 95 L/min)	≤ 70 Pa (arithmetic) + 50 Pa (no valve)
Flow rate	85 L/min	Varied - see above	85 L/min	Varied - see above	Varied - see above	40 L/min
Exhalation resistance - max pressure drop	≤ 243 Pa	≤ 300 Pa	≤ 250 Pa	≤ 300 Pa	≤ 300 Pa	≤ 70 Pa (arithmetic) + 50 Pa (no valve)
Flow rate	85 L/min	80 L/min	85 L/min	85 L/min	80 L/min	40 L/min
Exhalation valve leakage requirement	Leak rate ≤ 30 mL/min	N/A	Depressurization ≤ 0.1 Pa @ 250 sec	Leak rate ≤ 30 mL/min	Leak rate ≤ 30 mL/min	Depressurization ≤ 0.1 Pa @ 30 sec
Force applied	≤ 245 Pa	N/A	≤ 350 Pa	≤ 250 Pa	N/A	≤ 1,470 Pa
City clearance requirement	N/A	≤ 7%	≤ 7%	≤ 7%	≤ 7%	≤ 7%

*Japan JMAHLW-Notification 214 requires an Inward Leakage test rather than a TIL test.

¿Dónde usarlo?: Oficina, Escuela, Restaurante, Barberia

Cantidad x Empaque: 50

Características: Reutilizable para mas de un uso con filtro triple.

Código: ZP-0155

Color: Blanco

Filtros Tipo: N95

Procedencia: Importado

Se vende por: Paquetes de 50 unidades.

Usos: Multiproposito

Ajustable: Si

Capacidad: Filtra el 95% de los virus y bacterias mismo nivel de protección que NIOSH N95

Certificaciones: CE

Código de producto: ZP-0155

Con Banda: Si

Marca: YATO

Protección: Contra el polvo no tóxico, polvo de jardín, polvo de hogar o fibras que producen polvo y humo

Tipo: Respiradores y Filtros

KN95

MULTIPLE PROTECTION



FILTRATION EFFICIENCY
95%

-  Smong
-  Powder
-  Talgas
-  Second
-  Antrodor
-  Granules



Based on this comparison, it is reasonable to consider China KN95, AS/NZ P2, Korea 1st Class, and Japan DS FFRs as "equivalent" to US NIOSH N95 and European FFP2 respirators, for filtering non-oil-based particles such as those resulting from wildfires, PM2.5 air pollution, volcanic eruptions, or bioaerosols (e.g. viruses). However, prior to selecting a respirator, users should consult their local respiratory protection regulations and requirements or check with their local public health authorities for selection guidance.

Certification/ Class (Standard)	N95 (NIOSH-42C FR84)	FFP2 (EN 149-2001)	KN95 (GB2626-20 06)	P2 (AS/NZ 1716:2012)	Korea 1 st Class (KMOEL - 2017-64)	DS (Japan JMHLW- Notification 214, 2018)
Filter performance -- (must be ≥ 95% efficient)	≥ 95%	≥ 94%	≥ 95%	≥ 94%	≥ 94%	≥ 95%
Test agent	NaCl	NaCl and paraffin oil	NaCl	NaCl	NaCl and paraffin oil	NaCl
Flow rate	85 L/min	95 L/min	85 L/min	95 L/min	95 L/min	85 L/min
Total inward leakage (TIL)* -- tested on human subjects each performing exercises	N/A	≤ 8% leakage (arithmetic mean)	≤ 8% leakage (arithmetic mean)	≤ 8% leakage (individual and arithmetic mean)	≤ 8% leakage (arithmetic mean)	Inward Leakage measured and included in User Instructions
Inhalation resistance -- max pressure drop	≤ 343 Pa	≤ 70 Pa (at 30 L/min) ≤ 240 Pa (at 95 L/min) ≤ 500 Pa (clogging)	≤ 350 Pa	≤ 70 Pa (at 30 L/min) ≤ 240 Pa (at 95 L/min)	≤ 70 Pa (at 30 L/min) ≤ 240 Pa (at 95 L/min)	≤ 70 Pa (w/valve) ≤ 50 Pa (no valve)
Flow rate	85 L/min	Varied -- see above	85 L/min	Varied -- see above	Varied -- see above	40 L/min
Exhalation resistance -- max pressure drop	≤ 245 Pa	≤ 300 Pa	≤ 250 Pa	≤ 120 Pa	≤ 300 Pa	≤ 70 Pa (w/valve) ≤ 50 Pa (no valve)
Flow rate	85 L/min	160 L/min	85 L/min	85 L/min	160 L/min	40 L/min
Exhalation valve leakage requirement	Leak rate ≤ 30 mL/min	N/A	Depressurizatio n to 0 Pa ≥ 20 sec	Leak rate ≤ 30 mL/min	Leak rate ≤ 30 mL/min	visual inspection after 300 L /min for 30 sec
Force applied	-245 Pa	N/A	-180 Pa	-250 Pa	N/A	Depressurizatio n to 0 Pa ≥ 15 sec
CO ₂ clearance requirement	N/A	≤ 1%	≤ 1%	≤ 1%	≤ 1%	≤ 1%

*Japan JMHLW-Notification 214 requires an Inward Leakage test rather than a TIL test.